


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-2-P*

S.N.: *1356*

Date: *July 26 83*



- a) Output signal amplitude:  
*0 to +10 AMPS*
- b) Pulse width:  
*0 TO 15 NSEC*
- c) Rise time:  
*≤ 1 NSEC*
- d) Fall time:  
*≤ 2 NSEC*
- e) PRF:  
*0 to 20 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*≈ 7 AMP/DIV.  
5 NSEC/DIV  
10 KHZ.*

*+15 ± 1.0 V*

*250 mA (max)*

*[Signature]*