


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVL-AV-1-PN*

S.N.: *1367*

Date: *MAY 25 83*



- a) Output signal amplitude:  
*0 TO ±100 V*
- b) Pulse width:  
*0 TO 100 nSEC*
- c) Rise time:  
*≤ 1 nSEC*
- d) Fall time:  
*≤ 2 nSEC*
- e) PRF:  
*0 TO 5 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+15 ± 1.0 V*  
*350 mA (MAX)*

*100 V / DIV*  
*5.0 nSEC / DIV*  
*4.0 KHZ.*