


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-1-P-D*

S.N.: *1375.*

Date: *June 29 83*



a) Output signal amplitude:

*0 to +5V*

b) Pulse width:

*0 to 6 nsec*

c) Rise time:

*≤ 100 psec*

d) Fall time:

*≤ 135 psec*

e) PRF:

*0 to 25 MHz*

f) Jitter, stability:

*OK*

g) Prime power:

*10 volts/DIV.*

*1.0 nsec/DIV.*

*4.0 MHz.*

*+24 ± 1.0 V*

*400 mA (MAX)*