


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PMM-1-P-D*

S.N.: *1376*

Date: *JUNE 29 83*



- a) Output signal amplitude:
0 TO +5V
- b) Pulse width:
0 TO 6 NSEC
- c) Rise time:
< 100 PSEC
- d) Fall time:
< 135 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*10 VOLTS/DIV.
1.0 NSEC/DIV
4.0 MHz*

*+24 ± 1.0 V
400 mA (MAX)
*[Signature]**