


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-1-P-D*

S.N.: *1377*

Date: *JUNE 29 83*



*10 VOLTS/DIV.
1.0 NSEC/DIV.
4.0 MHz.*

- a) Output signal amplitude:
0 TO +5V
- b) Pulse width:
0 TO 6 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC.
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*+24 ± 1.0 V
400 mA (MAX)*

AK