


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVK-A-C-P*

S.N.: *1382*

Date: *JULY 28 83*



- a) Output signal amplitude:  
*0 to +15V*
- b) Pulse width:  
*2 to 100 nsec*
- c) Rise time:  
*≤ 100 psec*
- d) Fall time:  
*≤ 2 nsec*
- e) PRF:  
*0 to 1 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*10 VOLTS/DIV*  
*1.0 nsec/DIV*  
*400 kHz*

*110/220V*  
*50-60 Hz*