


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-2-P

S.N.: 1383

Date: Jul 28 83



- a) Output signal amplitude:
0 to +350 V
- b) Pulse width:
0 to 100 nsec
- c) Rise time:
≤ 1.5 nsec
- d) Fall time:
≤ 2 nsec
- e) PRF:
0 to 5 kHz
- f) Jitter, stability:
OK
- g) Prime power:

≈ 320 V/DIV
10 nsec/DIV
2 kHz

+15 ± 1.0 V
500 mA (max)