


PULSE GENERATOR

PERFORMANCE CHECK

Model: AV7-AV-15-C-05-M-N

S.N.: 1384

Date: AUG 29 83



5 VOLTS/DIV
1.0 NSEC/DIV
400 KHZ

a) Output signal amplitude:

0 TO -5V

b) Pulse width:

0 TO 4 NSEC

c) Rise time:

≤ 60 PSEC

d) Fall time:

≤ 135 PSEC

e) PRF:

0 TO 1 MHZ


f) Jitter, stability:

OK

g) Prime power:

110/220V

50-60 HZ



5 VOLTS/DIV
100 PSEC/DIV
400 KHZ

[Handwritten signature]