


PULSE GENERATOR

PERFORMANCE CHECK

Model: *APD-AU-1-N-TTL-PS*

S.N.: *1386*

Date: *AUG 29 83*



*10 VOLTS/DIV
1.0 NSEC/DIV
400 KHZ.*

- d) Output signal amplitude:
0 to -10V
- e) Pulse width:
0 TO 4 NSEC
- f) Rise time:
≤ 100 PSEC
- g) Fall time:
≤ 135 PSEC
- h) PRF:
0 TO 1.0 MHz
- i) Jitter, stability:
OK

g) Prime power:
*110/220V
50-60 HZ.*