


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AUM-2-N-M*

S.N.: *1392*

Date: *JUNE 27 83*



- a) Output signal amplitude:
0 TO 15 V
- b) Pulse width:
0 TO 2 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MHz.
- f) Jitter, stability:
OK
- g) Prime power:

10 VOLTS/DIV
1.0 NSEC/DIV
4.0 MHz.

- +24 ± 1.0 VOLTS,*
500 mA MAX.
- h) MONITOR OUTPUT:
0 TO 1.5 V
(20 db DOWN WRT
MAIN OUTPUT).