


PULSE GENERATOR

PERFORMANCE CHECK


Model: ANI-V-1703-P

S.N.: 1395

Date: SEPT 27 83



100 V/DIV
10 nSEC/DIV
4 KHz



100 V/DIV
1.0 nSEC/DIV
4 KHz

a) Output signal amplitude:

0 TO 7100V

b) Pulse width:

0 TO 100 nSEC

c) Rise time:

≤ 200 pSEC

d) Fall time:

~ 2 nSEC

e) PRF:

0 TO 5 KHz

f) Jitter, stability:

OK

g) Prime power:

+15 ± 1.0 V

275 mA (MAX)