


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVL-2A-C-W-P

S.N.: 1406

Date: OCT 4 83



- a) Output signal amplitude:  
0 TO +170 V
- b) Pulse width: \*  
0 TO 400 nSEC
- c) Rise time:  
L 1 nSEC
- d) Fall time:  
L 2 nSEC
- e) PRF:  
0 TO 5 KHz
- f) Jitter, stability:  
OK
- g) Prime power:

100 V/DIV.  
10 nSEC/DIV

4 KHz.  
( $R_L = 50 \Omega$ )

110/220 V  
50-60 Hz.

\* REQUIRES ADDITIONAL  
EXTERNAL DELAY  
LINE FOR PULSE  
100-400 nSEC.