

PULSE GENERATOR

PERFORMANCE CHECK

Model:


AVP-AV-15-C-P

S.N.:


1407

Date:

AUG 29 83



5 V/DIV
1.0 nsec/DIV
40. kHz.



100 psec/DIV
5 VOLTS/DIV
400 kHz.

a) Output signal amplitude:

0 TO +5V

b) Pulse width:

0 TO 4 nsec

c) Rise time:

≤ 60 psec

d) Fall time:

≤ 135 psec

e) PRF:

0 TO 1 MHz

f) Jitter, stability:

OK

g) Prime power:

110/220V,
50-60 Hz.