


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-TN-1*
S.N.: *1415*
Date: *JUNE 23 1963*



*100 V / DIV.
10 nSEC / DIV
2 kHz*

- a) Output signal amplitude:
- 230 V (+30 V OFFSET)
- b) Pulse width:
0 TO 100 nSEC
- c) Rise time:
≤ 2 nSEC
- d) Fall time:
≤ 3 nSEC
- e) PRF:
0 TO 5 kHz
- f) Jitter, stability:
OK
- g) Prime power:

*110 / 220 V
50 - 60 Hz
TH: ≤ 100 nSEC*