


PULSE GENERATOR
PERFORMANCE CHECK

Model: *A20-1-P*

S.N.: *1418*

Date: *MAY 29 83*



*64 AMPS / DIV.
10 NSEC / DIV
1 KHZ.*

- a) Output signal amplitude:
25 TO 100 AMPS (A)
- b) Pulse width:
25 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1.0 V

190 mA MAX