


PULSE GENERATOR
PERFORMANCE CHECK

Model: 1720-1-P
S.N.: 1419
Date: MAY 29 83



- a) Output signal amplitude: +25 to +100 AMPS
- b) Pulse width: 25 nSEC
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: 0 to 1 KHz
- f) Jitter, stability: OK
- g) Prime power: OK

64 AMPS/DIV.

10 nSEC/DIV.

1 KHz.

+15 ± 1.0V

100 mA (mA)