


PULSE GENERATOR

PERFORMANCE CHECK

Model: AV0-3B-17-P5

S.N.: 1430

Date: AUG 11 1983



- a) Output signal amplitude:
+10 to +30 AMPS
- b) Pulse width:
≤ 1.5 NSH
- c) Rise time: (HALF POWER)
NA
- d) Fall time:
NA
- e) PRF:
0 to 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

≈ 20 AMPS / DIV

2.0 NSH / DIV

10 KHz.

110/220V

50-60 Hz.