


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVMR-2-PW-P*

S.N.: *1441*

Date: *JUNE 13 83*



- a) Output signal amplitude:  
*0 TO +20V*
- b) Pulse width:  
*20 TO 200 NSEC*
- c) Rise time:  
*≤ 3 NSEC*
- d) Fall time:  
*≤ 3 NSEC*
- e) PRF:  
*0 TO 10 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*10 VOLTS/DIV*  
*10 NSEC/DIV*  
*4 MHz*

*+24V ± 1.0V*  
*200 mA*  
b) *DUTY CYCLE:*  
*20% MAX.*

*AY*