


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-TN-1
S.N.: 1446
Date: Jan 23 83



100 VOLTS/DIV
10 NS/DIV
2 KHz

- a) Output signal amplitude:
-230 V (+30V OFFSET)
- b) Pulse width:
0 TO 100 NS
- c) Rise time:
< 2 NS
- d) Fall time:
< 3 NS
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
110/220V
50-60 Hz

tp: < 100 NS
[Signature]