


PULSE GENERATOR
PERFORMANCE CHECK

Model: PVL-TN-1

S.N.: 1448

Date: JUNE 23 83



100 VOLTS/DIV
10 NS/DIV
2 KHz.

- a) Output signal amplitude:
-230 V (13V OFFSET)
 - b) Pulse width:
0 TO 100 NS
 - c) Rise time:
≤ 2 NS
 - d) Fall time:
≤ 3 NS
 - e) PRF:
0 TO 5 KHz.
 - f) Jitter, stability:
OK
 - g) Prime power:
110/220 V
50-60 Hz.
- Ep: ≤ 100 NS

[Signature]