


PULSE GENERATOR
PERFORMANCE CHECK

Model: ~~AA~~ AVL-TN-1

S.N.: 1649

Date: ~~June 15~~ 23 83



100 V/DIV
10 nSEC/DIV
2 kHz

- a) Output signal amplitude:
- 230 V (+30 V OFFSET)
 - b) Pulse width:
0 TO 100 nSEC
 - c) Rise time:
< 2 nSEC
 - d) Fall time:
< 3 nSEC
 - e) PRF:
0 TO 5 kHz
 - f) Jitter, stability:
OK
 - g) Prime power:
110/220V
50-60 Hz
tp: < 100 nSEC
- [Signature]*