


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO-1*

S.N.: *1454*

Date: *AUG 30 83*



- a) Output signal amplitude:
25 40 100 AMPS
- b) Pulse width:
25 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

*= 60 VAMPS/DIV.
10 NSEC/DIV
1 KHZ*

*+15 ± 1.0 V
160 mA (MAX)*