


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVP-AV-2-P*

S.N.: *1456*

Date: *04 14 83*



a) Output signal amplitude:

*0 to +10V*

b) Pulse width:

*5 to 50 nsec*

c) Rise time:

*≤ 100 psec*

d) Fall time:

*≤ 135 psec*

e) PRF:

*0 to 50 kHz*

f) Jitter, stability:

*OK*

g) Prime power:

*+15 ± 1.0 V*

*275 mA (MAX)*

*10 VOLTS/DIV*

*1.0 nsec/DIV*

*30 kHz*