


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVM-2-N-PS

S.N.: 1458

Date: DEC. 22, 1983



a) Output signal amplitude:

0 TO 15V

b) Pulse width:

0 TO 2 NS

c) Rise time:

≤ 100 PSEC

d) Fall time:

≤ 135 PSEC

e) PRF:

0 TO 25 MHz

f) Jitter, stability:

OK

g) Prime power:

110/220V

50-60Hz

[Signature]

10 VOLTS/DIV

100 NS/DIV

4.0 MHz