


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-2-N-PS*

S.N.: *1459*

Date: *DEC 22, 1983*



*10 VOLTS/DIV
1.0 NSEC/DIV
4.0 MHz*

a) Output signal amplitude:

0 TO 15V

b) Pulse width:

0 TO 2 NSEC

c) Rise time:

≤ 100 PSEC

d) Fall time:

≤ 135 PSEC

e) PRF:

0 TO 25 MHz

f) Jitter, stability:

OK

g) Prime power:

110/220V

50-60Hz