


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AV6-1-P

S.N.: 1460

Date: AUG 29 83



- a) Output signal amplitude:
0 TO +250 VOLTS
- b) Pulse width:
≤ 4 NSEC
- c) Jitter, stability:
OK
- d) Prime power:
+15 ± 1.0 V
350 mA (max)

[Handwritten signature]

100 VOLTS/DIV
2.0 NSEC/DIV
40 KHz
(Resonance)