





PULSE GENERATOR
PERFORMANCE CHECK

Model: *NI-V-14V3-C-PN-03-M*

S.N.: *1463*

Date: *OCT 11 83*



- a) Output signal amplitude:
0 to ±100V
- b) Pulse width:
0 TO 100 NS ± 2
- c) Rise time:
≤ 200 PSEC (P)
≤ 300 PSEC (N)
- d) Fall time:
≤ 2 NS ± 2
- e) PRF:
0 to 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

110/220V
50-60 Hz.