


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-17U-1-PSP-TTL*

S.N.: *1464*

Date: *AUG 29 83*



- a) Output signal amplitude:
0 TO +10V
- b) Pulse width:
0 TO 4 nSEC
- c) Rise time:
≤ 100 pSEC
- d) Fall time:
≤ 135 pSEC
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*10 VOLTS/DIV
100 nSEC/DIV
400 kHz*

*120/240 V
50-60 Hz*

*UNIT SHIPPED
WITH VOLTAGE
SELECTOR SWITCH IN
THE 240 VOLT
POSITION.*
[Signature]