

PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVN-1-P*

S.N.: *1474*

Date: *AUG 30 83*

a) Output signal amplitude:

*0 TO +5V*

b) Pulse width:

*0.2 TO 1.0 nSEC*

c) Rise time:

*≤ 100 PSEC*

d) Fall time:

*≤ 135 PSEC*

e) PRF:

*50 - 260 MHz*


f) Jitter, stability:

*OK*

g) Prime power:

*+24 ± 1.0 V*

*500 mA (MAX)*



*5 VOLT/Div  
1.0 nSEC/Div  
150 MHz*