


PULSE GENERATOR

PERFORMANCE CHECK

Model: AV0-1

S.N.: 1475

Date: Aug. 30 83



- a) Output signal amplitude: 25 TO 100 AMPS
- b) Pulse width: 26 NSEC
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: 0 TO 1 KHZ
- f) Jitter, stability: OK
- g) Prime power:

60 AMPS / DIV
10 NSEC / DIV
1 KHZ

+15 ± 1.0 V
160 mA (MAX)