

PULSE GENERATOR

PERFORMANCE CHECK

Model:


AVI-V-F

S.N.:

1478

Date:

OCT 14 83



a) Output signal amplitude:

0 TO +50V

b) Pulse width:

0 TO 100 nSEC

c) Rise time:

≤ 300 pSEC

d) Fall time:

≤ 2 nSEC

e) PRF:

0 TO 20 kHz

f) Jitter, stability:

OK

g) Prime power:

+15 ± 1.0 V

250 mA (MAX)

32 VOLTS/DIV

5 nSEC/DIV

10 kHz