


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-P
S.N.: 1487
Date: OCT 14 83



- a) Output signal amplitude:
0 TO +30V
- b) Pulse width:
≤ 1.0 nsec
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

3.2 VOLTS/DIV
1.0 nsec/DIV
400 kHz

+15 ± 1.0 V
150 mA
(MAX)

A handwritten signature or set of initials, possibly 'JA', written in dark ink at the bottom right of the page.