


PULSE GENERATOR
PERFORMANCE CHECK

Model: AV14-P
S.N.: 1488
Date: OCT 14 83

- a) Output signal amplitude:
0 TO +15.0 V
- b) Pulse width:
~ 1.0 nsec
- c) Rise time:
N/A
- d) Fall time:
N/A
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:



32 VOLTS/DIV
1.0 nsec/DIV
400 kHz

+15 ± 1.0 V

150 nH

(N/A)

A handwritten signature or set of initials, possibly 'JH', written in dark ink.