


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVI-P*

S.N.: *1490*

Date: *SEPT 26 83*



- b) Output signal amplitude:
0 TO +40V
- c) Pulse width:
2 TO 100 nsec
- d) Rise time:
< 500 psec
- e) Fall time:
< 2 nsec
- f) PRF:
800 PLS 1Hz
- g) Jitter, stability:
OK
- h) Prime power:
+15 ± 1.0 V
150 mA (MAX)

1.0 nsec/div
32 VOLTS/div
40 kHz