


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-2-C-R-PN-M-EX-EW-ED*

S.N.: *1496*

Date: *OCT 6 83*



- a) Output signal amplitude:
0 to ±15 UNITS
- b) Pulse width:
0 TO 2 nsec.
- c) Rise time:
≤ 100 psec
- d) Fall time:
≈ 135 psec
- e) PRF:
0 TO 25 kHz
- f) Jitter, stability:
0%
- g) Prime power:

*10 UNITS/DIV.
1.0 nsec/DIV
5 kHz*

*110/220V,
50-60 Hz.*

