

PULSE GENERATOR

PERFORMANCE CHECK

Model:


PT2-1

S.N.:

1498

Date:

1964 30 83



a) Output signal amplitude:

25 TO 100 AMPS

b) Pulse width:

26 NSCE

c) Rise time:

N/A

d) Fall time:

N/A

e) PRF:

0 TO 1 KHz

f) Jitter, stability:

OK.

g) Prime power:

$\pm 15 \pm 1.0V$

150 mA (MAX)

60 AMPS/PLU

10 NSCE/PLU

1 KHz