


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVP-AV-1-05-MM-P

S.N.: 1500

Date: MAR 24 83



- a) Output signal amplitude:
0 TO +10V
- b) Pulse width:
0 TO 4 NS
- c) Rise time:
≤ 100 psec
- d) Fall time:
≤ 133 psec
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

10 VOLTS/DIV

1.0 nsec/DIV

400 KHz

+15 ± 1.0V

250 mA (MAX)