


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PUL-MV-PS-N-M*

S.N.: *1502*

Date: *NOV 1 83*



- a) Output signal amplitude:
0 TO -250V
- b) Pulse width:
0 TO 100 NSEC
- c) Rise time:
≤ 5 NSEC
- d) Fall time:
≤ 5 NSEC
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

*100 VOLTS/DIV.
10 NSEC/DIV.
2 KHZ.*

*110/220V,
50-60 HZ.*