


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AMP-AU-1-N'*  
S.N.: *1505*  
Date: *AUG 24 83*



- a) Output signal amplitude:  
*0 TO -10V*
- b) Pulse width:  
*0 TO 4 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 1 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*10 VOLTS/DIV.*  
*1.0 NSEC/DIV*  
*400 KHz*

*+15 ± 1.0 V*  
*300 mA (MAX)*