


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVM-2-P

S.N.: 1507

Date: Sept. 30/83



10 V/DIV

1.0 NSEC/DIV

4.0 MHz

a) Output signal amplitude:

0 to +15V

b) Pulse width:

0 to 2 nsec

c) Rise time:

≤ 100 psec

d) Fall time:

≤ 135 psec

e) PRF:

0 to 25 MHz

f) Jitter, stability:

OK

g) Prime power:

+24 \pm 1.0 V

530 mA (max)