


PULSE GENERATOR

PERFORMANCE CHECK

Model: *ANM-2-PS-P*

S.N.: *1510*

Date: *NOV 18, 1983*



a) Output signal amplitude:

*0 TO +15V...*

b) Pulse width:

*0 TO 2.502*

c) Rise time:

*≤ 100 PSEC*

d) Fall time:

*≤ 135 PSEC*

e) PRF:

*0 TO 25 kHz*

f) Jitter, stability:

*OK*

g) Prime power:

*110/220V*

*50-60 Hz.*

*10 VOLTS/DIV*

*100 NSEC/DIV*

*4.0 MHz.*