


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVN-1-N*
S.N.: *1512*
Date: *AUG 30 83*



- a) Output signal amplitude:
0 TO \approx -10V
- b) Pulse width:
0.2 TO \approx 2 nSEC
- c) Rise time: (*AT 100 MHz*)
 \leq 100 pSEC
- d) Fall time:
 \leq 135 pSEC
- e) PRF:
SD - 260 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*10 VOLTS/DIV
1.0 nSEC/DIV
100 MHz*

*+ 24 \pm 1.0 V
500 mA (MAX)*