

PULSE GENERATOR
PERFORMANCE CHECK

Model:


AVK-V-P

S.N.:

1514

Date:

OCT 26 1963



a) Output signal amplitude:

$\approx +5V$

b) Pulse width:

0 TO 50 NSEC

c) Rise time:

$\leq 100 \mu\text{SEC}$

d) Fall time:

$\leq 2 \text{ NSEC}$

e) PRF:

0 TO 50 KHZ

f) Jitter, stability:

OK

g) Prime power:

10 VOLTS/DIV
1.0 USEC/DIV
40 KHZ

$\pm 15 \pm 1.0 \text{ VOLT}$
250 MA (MAX)

[Signature]