


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVP-AV-1-PS-OS-PN

S.N.: 1520

Date: DEC 22, 1983



- a) Output signal amplitude:
0 TO $\pm 10V$
- b) Pulse width:
0 TO 4 NSEC
- c) Rise time:
 ≤ 100 PSEC
- d) Fall time:
 ≤ 135 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

10 VOLTS/DIV
1.0 NSEC/DIV
400 KHz

110/220V
50-60Hz