


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PK-AU-05-TR-N*

S.N.: *1523*

Date: *OCT 26 83*



- a) Output signal amplitude: *0 to 15V*
- b) Pulse width: *0 to 50 nsec*
- c) Rise time: *≤ 100 psec*
- d) Fall time: *≤ 2 nsec.*
- e) PRF: *0 to 50 kHz*
- f) Jitter, stability: *OK*
- g) Prime power: *OK*

*1.0 nsec/div
10 volts/div
40 kHz.*

*+15 ± 1.0 V
250 mA (max)*