

PULSE GENERATOR

PERFORMANCE CHECK

Model:


AVK-A-C-P

S.N.:

1528

Date:

DEC 14 83



a) Output signal amplitude:

0 TO +15V.

b) Pulse width:

4 TO 100 nSEC

c) Rise time:

≤ 100 pSEC

d) Fall time:

≤ 2 nSEC

e) PRF:

SEE FIG 1, 2

f) Jitter, stability:

OK

g) Prime power:

110/220V

50-60 Hz.

10 VOLT/DIV

2 nSEC/DIV

100 kHz.

*[Signature]*