


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-2-N*

S.N.: *1530*

Date: *SEPT 15 83*



a) Output signal amplitude:

*0 TO 15V*

b) Pulse width:

*0 TO 2 nSEC*

c) Rise time:

*≤ 100 PSEC*

d) Fall time:

*≈ 135 PSEC*

e) PRF:

*0 TO 25 MHz*

f) Jitter, stability:

*OK*

g) Prime power:

*+24 ± 1.0 V*

*500 mA (max)*

*1.0 nSEC/DIV*

*10 V/DIV*

*4.0 MHz*

*[Signature]*