


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVP-AV-1AV3-C-PN-05-M*

S.N.: *1539*

Date: *FEB 13 84*



*32 VOLTS DIV
100 NSEC/DIV
100 KHZ.*

- a) Output signal amplitude:
0 TO ±40V
- b) Pulse width:
0 TO 2 nsec
- c) Rise time:
≤ 150 psec
- d) Fall time:
≤ 250 psec
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
0%
- g) Prime power:
110/220V

50 .60 Hz
[Signature]