


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-N
S.N.: 1548
Date: DEC 23 83



- a) Output signal amplitude:
0 TO -30V ...
- b) Pulse width:
≤ 1 nsec
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

32 VOLTS/DIV
1.0 nsec/DIV
400 kHz

+15 ± 1.0 V
150 mA (MAX)
My