


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVN-2-PS-MM-1²

S.N.: 1566

Date: NOV 29 F.3



- a) Output signal amplitude:
0 TO +2V
- b) Pulse width:
0.2 TO 1.5 NSER
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC.
- e) PRF:
50 TO 250 MHz.
- f) Jitter, stability:
OK
- g) Prime power:

1.0 NSER/DIV
5.0 VOLTS/DIV.

110/220V
50-60 Hz
JH